

Solid immersion microscopy readily and inexpensively enables 12 nm resolution on plunge-frozen cells

Lin Wang^{a,1}, Benji Bateman^{a,1}, Laura C. Zanetti-Domingues^{a,1}, Amy N. Moores^a, Sam Astbury^a, Christopher Spindloe^a, Michele C. Darrow^b, Maria Romano^{c,d}, Sarah R. Needham^a, Konstantinos Beis^{c,d}, Daniel J. Rolfe^a, David T. Clarke^a and Marisa L. Martin-Fernandez^{a,2}

^aScience and Technology Facilities Council, Central Laser Facility, Rutherford Appleton Laboratory, Didcot OX11 0QX, United Kingdom

^bDiamond Light Source, Harwell Science and Innovation Campus, Didcot OX11 0DE, United Kingdom

^cDepartment of Life Sciences, Imperial College London, London SW7 2AZ, United Kingdom

^dResearch Complex at Harwell, Rutherford Appleton Laboratory, Didcot OX11 0FA, United Kingdom

¹These authors contributed equally to the work

²To whom correspondence should be addressed. Email: marisa.martin-fernandez@stfc.ac.uk

Keywords: Cryo-fluorescence microscopy, solid immersion lens, STORM, super-resolution, cell imaging

Running title: Low-cost/tech STORM for 12 nm resolution

Abstract

Super-resolution fluorescence microscopy achieves 20-30 nm resolution by using liquid-immersion objectives to optimize light collection and chemical sample fixation to minimize image blurring. It is known that fluorophore brightness increases substantially under cryogenic conditions and that cryo-fixation is far superior in preserving ultrastructure. However, cryogenic conditions have not been exploited to improve resolution or sample quality because liquid immersion media freezes at the objective, losing its optical properties. Here, simply by replacing the immersion fluid with a low-cost super-hemispherical solid immersion lens (*superSIL*), we effortlessly achieve <8 nm localisation precision and 12 nm resolution under cryogenic conditions in a low-cost, low-tech system. This is to our knowledge the best resolution yet attained in biological samples. Furthermore, we demonstrate multicolour imaging and show that the inexpensive setup outperforms 10-fold more costly super-resolution microscopes. By also removing the barrier to total internal reflection fluorescence imaging of mammalian cells under cryogenic conditions, *superSIL* microscopy delivers a straightforward route to achieve unmatched nanoscale resolution on both bacterial and mammalian cell samples, which any laboratory can effortlessly and inexpensively implement.

Introduction

Super-resolution fluorescence microscopy (FM) techniques, such as structured illumination microscopy (SIM) [1], stimulated emission depletion (STED) microscopy [2], and single molecule localization microscopy (SMLM) (including stochastic optical reconstruction microscopy (STORM) [3], and photoactivated localization microscopy (PALM) [4, 5]), have provided nanoscale insights into the functioning of cellular components for over a decade. However, technical complexity and high costs have so far prevented their general application to the biomedical field outside specialist laboratories. Furthermore, nanoscale resolution typically requires chemical fixation to prevent blurring. In electron microscopy (EM), vitrification through rapid freezing has proven more effective at preserving ultrastructure [6], yet there are few examples in the literature of its use in super-resolution. This is despite the resolution improvement that can be expected from the increased photon budget available under cryogenic conditions. To achieve super-resolution in cryogenic conditions is a timely goal to realize the promise of correlative light and electron microscopy (CLEM) in biology [7], so far limited by the resolution mismatch between cryo-EM and cryo-FM, largely restricting the optical element to identification of regions of interest.

The resolution of SMLM techniques, like STORM, depends on the precision by which individual molecules can be localized in the sample. This in turn depends on the number of photons emitted, which increases under cryogenic conditions, and the number collected by the objective lens, which is determined by its numerical aperture (NA) [8, 9]. NAs > 1 and preferably > 1.4 are required to achieve high resolution, but these need immersion fluids to couple the sample to the objective. Because liquid media freeze, dry objectives (NA ≤ 0.9) are mostly used under cryogenic conditions, with the disadvantages of lower resolution and limited photon collecting ability. The larger photon yield of common fluorophores under cryogenic conditions nevertheless allowed Kaufman *et al* [10] to achieve ~ 125 nm resolution on green fluorescence protein (GFP)-stained samples. Liu *et al* [11] attained ~ 74 nm resolution using DRONPA, a protein 2.5 times brighter than GFP. Recent work [12, 13] reported highly specialized cryo-STORM systems functioning at liquid Helium temperature, which yielded an astonishing \sim Angstrom localization precision on isolated molecules. Other set-ups relied on custom-built stages to incorporate cryofluids; Nahmani *et al* [14] achieved ~ 35 nm resolution using a water-immersion objective (NA = ~ 1.2). A similar approach was taken by Faoro *et al* [15], but this set-up was not applied to super-resolution microscopy. However, the complex machinery required has rendered cryo-STORM so far largely inapplicable to general cell biology. Furthermore, lenses with NAs < 1.4 , which cannot create evanescent fields, are incompatible with the TIRF illumination required for super-resolution imaging of mammalian cells [16].

Super-hemispherical solid immersion lenses (*superSILs*) are truncated balls made of solid materials of high refractive index and fill the gap between the objective and the sample, eliminating the requirement for coupling fluids [17-19]. By using a *superSIL* to couple the sample to a dry objective, the effective NA of the latter is enhanced up to the value of the refractive index of the *superSIL*. We demonstrated the high-resolution of *superSIL*-based FM under low signal-to-noise-ratio conditions at room temperature [20, 21]. Proof-of-concept experiments were also carried out in STED [22, 23] and SIM [24]. However, *superSILs* are particularly suited to STORM techniques because of the critical dependence on high photon collection efficiency. Indeed, it was previously shown that *superSILs* increased photon collection efficiency by ~6-fold [25].

Here, we describe a low-cost, low-complexity STORM setup that exploits solid immersion technology to effortlessly achieve an unprecedented 12 nm resolution. We solved the problem of how to exploit the increased fluorophore brightness under cryogenic conditions and removed the barrier to TIRF imaging of cryo-vitrified samples and outperformed a much more expensive state-of-the-art system at room temperature too. We expect these properties to make *superSIL* microscopy the method of choice for the general implementation of nanoscale imaging in non-specialist laboratories.

Results and Discussion

Our aim was to find a low-tech solution that could be effortlessly and inexpensively implemented in conventional low-cost optical microscopes and had the flexibility to exploit cryogenic conditions when improving resolution and/or avoiding artefacts from chemical fixation were necessary (**Fig. 1**). The optics to deliver and collect light consist of a 0.55 NA 100× dry objective (Mitutoyo 100× Plan Apo SL Infinity Corrected) and a 1 mm diameter *superSIL* made of cubic zirconia (effective NA = 2.17). An achromatic doublet lens with a focal length of 200 mm was used as tube lens before the EMCCD detector camera (Andor, iXon+ DU-897). The *superSIL* was mounted into the central hole of a platinum disk using a thermally conductive cryo-adhesive (Loctite Stycast 2850 FT epoxy encapsulant) (**Fig. 1A**). *SuperSIL* assemblies are robust and inexpensive (~\$20), and can be reused by cleaning the surfaces with trypsin, common solvents, and/or piranha solution.

The *superSIL* assembly was designed to be compatible with the standard EM grid-holder of an off-the-shelf liquid nitrogen (LN₂)-cooled cryo-stage (Linkam, CMS-196). The samples were adhered to the flat surface of the *superSILs* and plunge-frozen together in liquid ethane. The convex surfaces of the *superSILs* faced upwards to the objective, and the flat surfaces, onto which the samples are adhered, faced downwards (**Fig. 1B**). *SuperSIL* assemblies thus play at any temperature, including cryogenic conditions, the role played by liquid immersion media at room temperature. The *superSIL* assemblies were mounted on top of the brass bridge in the cryo-stage and translated in XY directions for fine positioning. The cryo-stage is self-contained with a built-in LN₂ reservoir, and can be easily

integrated into any conventional upright fluorescence microscope. Frozen samples were kept at $-196\text{ }^{\circ}\text{C}$ by the LN_2 vapour surrounding the brass bridge of the cryo-stage.




Fig. 1| Schematic of *superSIL* microscope. (A) Side and top view of *superSIL* assembly. The side view shows the location of the sample plane (aplanatic surface). The photo of an assembly is at the bottom of the panel. (B) Schematic of key components in the microscope. An upright microscope configuration was employed. The blue shading illustrates the Köhler illumination laser beam, the green shading indicates the fluorescence emission. (For more info, see **Supporting Information, Fig. S1-S3**).

The first step to characterize the performance of the *superSIL* microscope was to measure the size of the point spread function (PSF) of the microscope, i.e. the image of a point object, at room temperature, for which we did not cool the stage. We acquired 500 photoluminescence images of the sparse cubic zirconia defect spots present on the flat surface of the *superSIL*s (**Fig. S4**), and measured their full width at half maximum (FWHM). These intrinsic defects are single point emitters [26], and are located at the aplanatic plane of the *superSIL*, which is the position with the fewest optical aberrations [20]. Thus, their images give an estimation of the best lateral resolution [27]. Representative images and the statistical results show a resolution of $153 \pm 15\text{ nm}$ (**Fig. 2A**), in agreement with that previously reported by a similar system [21]. The $\sim 25\%$ discrepancy between experimental and theoretical values (118 nm predicted using the Houston criterion [28]) arises from the failure to collect rays propagating at angles approaching 90° [20].

We next compared the performance of the *superSIL* microscope with that of a state-of-the-art, off-the-shelf STORM microscope (ZEISS Elyra equipped with an Alpha Plan-Apochromat $100\times/1.46\text{ NA}$

oil immersion DIC objective), which we hereafter refer to as specialist STORM system. As a test sample, we imaged live *E. coli* cells expressing the antibacterial peptide ABC transporter McjD fused with enhanced GFP (EGFP) [29-32]. To reach quasi-equilibrium between the fluorescent on and off states of the EGFP molecules, power densities of 0.349 kW/cm^2 and 1.1 kW/cm^2 ($\lambda = 488 \text{ nm}$) were delivered at the sample plane in the *superSIL* and specialist systems, respectively. As a control we imaged samples of *E. coli* expressing non-fluorescent McjD transporter with no EGFP (**Fig. S5**).

Typical FM images of EGFP-stained bacterial features from the *superSIL* and specialist microscopes are shown in **Fig. 2B**. Of note, the combination of a *superSIL* with a $100\times$ dry objective lens results in a magnification of $471\times$, which is an enhancement of nearly 5 times over a conventional ZEISS $100\times$ oil immersion objective. **Fig. 2C** shows representative STORM images and the corresponding molecular localization precision histograms. These reveal a 2-fold enhancement in localization precision using the *superSIL* STORM system ($15.4 \pm 4.1 \text{ nm}$) compared with the specialist STORM ($31.3 \pm 9.3 \text{ nm}$). *SuperSIL* STORM also reported sub-cellular structures that the specialist STORM system failed to resolve because overlapping blinking events were not accounted for during the image reconstruction in order to achieve high accuracy of localizations. This illustrates the benefit of the increased magnification provided by the *superSIL* system in avoiding PSF crowding from high-density labelling scenarios, in particular when imaging smaller cells (e.g. bacteria, yeast, etc), in which it is difficult to resolve structural details [33].

Sample structure and labelling density can have an effect on resolution [34]. To ascertain the best molecular localization precision we could obtain with the specialist system, we imaged standard calibration samples of DNA origami [35]. These custom nanorulers (GattaQuant) carried ATTO 647N dye molecules, ~ 20 times brighter than EGFP [36]. As shown in **Fig. 2D**, results revealed a molecular localization precision of $17.6 \pm 8.1 \text{ nm}$, in line with the microscope manufacturer's specifications. Interestingly, the best localization precision we could extract from the specialist system using ATTO 647N was no better than the $15.4 \pm 4.1 \text{ nm}$ returned by the *superSIL* system using the much dimmer EGFP fluorophore. We therefore conclude that the *superSIL* microscope outperforms the specialist STORM system. Moreover, the *superSIL* microscope costs ~ 20 times less than the specialist STORM system.

We next characterized the performance of the *superSIL* system under cryogenic conditions (**Fig. 3**). We first verified that the freezing procedure does not alter the optical properties of the *superSILs* (**Fig. S6**). As shown in **Fig. 3A** and **3C**, the statistical results of PSF measurements from cubic zirconia defect spots confirmed that the resolution of the *superSIL* FM system ($153 \pm 14 \text{ nm}$) is not perturbed by being plunge-frozen together with the sample.




Fig. 2| SuperSIL resolution performance at room temperature. (A) Image of a point object (*top*). Bar: 0.5 μm . The box chart (*bottom*) of the full width half maximum (FWHM) measurements of images from point emitters. The boxes show the 25th and 75th percentiles of the data points; the whiskers the standard deviation. (B) Comparison of wide-field images of live McjD-EGFP in *E.coli* cells. The *superSIL* image (*left*) has 471 \times magnification, while the specialist microscope (*right*) has 100 \times magnification. For comparison the inset contains a 4.71 \times scaled-up image of a cell. Bar: 5 μm , and 1 μm in the inset. (C) STORM images (*top*) and precision histograms (*bottom*) of live McjD-EGFP *E.coli* cells in *superSIL* STORM (*left*) and specialist STORM (*right*). The insets show wide-field fluorescence images of the cells. Bar: 1 μm . (D) Resolution evaluation of STORM imaging in the specialist microscope. *Left*: Schematic of DNA origami nanorulers labelled with ATTO 647N dye molecules. *Right*: image of a field of nanorulers obtained from wide-field (*top left*) and STORM (*top right*). Bar: 200 nm. The STORM image of a nanoruler is shown in the inset, indicated by the white-border box. Bar: 20 nm. *Bottom left*: Localization precision histogram from the STORM image of the nanorulers. *Bottom right*: Line profile of the cross-section of the nanoruler (*dashed line*) in the magnified inset, and its Gaussian fit (*continuous line*).

For comparison, we evaluated the size of the PSF of a cryo-compatible FEI CorrSight fluorescence microscope equipped with a standard 0.9 NA 40x dry objective (ZEISS EC Plan-Neofluar 40x/0.9 Pol M27) and its auxiliary cryo-chamber, hereafter referred to as specialist cryo-STORM system. A representative image and the statistical results of the PSF measurement using sub-diffraction limited 100 nm fluorescent beads (ThermoFisher, TetraSpeck) are shown in **Fig. 3B** and **3C**. These results report a mean PSF size of (328 ± 86) nm, which is consistent with the theoretical value of 290 nm at the wavelength of 515 nm, according to Houston criterion [28]. Thus, the lateral resolution of *superSIL* FM is about 2-fold better than that of the specialist system.

We next characterized the resolution achievable using STORM. For this we imaged in LN₂ vapour, plunged-frozen *E. coli* cells expressing the ATP-binding cassette (ABC) transporter PH1735, a putative multidrug transporter [37], fused to EGFP, which displayed robust single molecule blinking behaviour at cryogenic temperature (**Fig. S7A, S7B**). Representative wide-field and STORM images are shown in **Fig. 3D** and **3E**. The localization precision of 7.7 ± 3.2 nm (**Fig. 3F**) obtained with cryo-*superSIL* is 4.5-fold better than that of the specialist cryo-STORM system (35.7 ± 9.4 nm) (**Fig. S7C, S7D**). The latter value is consistent to that previously achieved in comparable conditions [10].

To reach quasi-equilibrium between the fluorescent on and off states of the EGFP molecules, we delivered a 1.1 kW/cm^2 laser power at 488 nm wavelength to the sample plane of both the *superSIL* and specialist cryo-STORM systems. Interestingly, the substantially larger photon delivery and collection efficiency of the *superSIL*s allowed us to use a 5.55 mW laser beam, whilst for specialist cryo-STORM imaging we required a 60 mW 488 nm laser beam. Thus, cryo-*superSIL* not only unlocks the potential to perform nanoscale STORM under cryogenic conditions, but makes this compatible with the use of low-power, low-cost lasers.

To illustrate the resolution of the cryo-*superSIL* system, we expanded the area within the red box in **Fig. 3E** (**Fig. 3G**), and then further enhanced in the latter the area within the orange box. The resulting image (**Fig. 3H**) shows single PH1735-GFP molecules embedded in the membrane together with larger features. The profiles of fluorescence intensity across the areas marked in **Fig. 3H** reveal two features separated by 65 nm (**Fig. 3I**), two features separated by ~ 16 nm (**Fig. 3J**), and two pairs of molecules separated by 12 nm and 15 nm (**Fig. 3K**). PH1735 is a homodimeric ABC transporter (thus, fused with two EGFP molecules), and these features probably belong to a full transporter. Taking into account the linker length between the PH1735 and EGFP, the measured separation of 12-17 nm probably represents a nucleotide free inward-open transporter, which is consistent with distances measured on other ABC transporters such as McjD [31]. To our knowledge, the combination of the <8 nm localization precision and 12 nm resolution here demonstrated is the highest to date on cells. We

objectively verified the resolution obtainable in cryo-*superSIL* STORM by imaging ATTO 647N-based DNA origami nanorulers (**Fig. S8**).




Fig. 3| *SuperSIL* resolution performance under cryogenic conditions. Representative images of sub-diffraction limited objects from (A) the *superSIL* microscope and (B) the specialist system. Bar: 0.5 μm . (C) Box charts of FWHM of the PSFs for the *superSIL* and specialist systems showing the 25th and 75th percentiles of the data points. The whiskers show the standard deviation. (D) ATP-binding cassette (ABC) transporter protein PH1735 fused with EGFP in *E. coli* cells imaged in wide-field *superSIL* FM mode and (E) *superSIL* STORM mode. (F) Localization precision histogram from the image in (E). (G) The enlarged image of the region of the cell indicated by the red dashed box in (E). (H) The further enlarged image of the small region in the cell indicated by the orange dashed box in (G). (I) Line profile of the cross-section of two PH1735 protein clusters indicated by the cyan dashed lines in (H). (J) Line profile of the cross-section of two PH1735 protein clusters closer to each other indicated by the red dashed line in (H). (K) Line profiles of the cross-section of two adjacent single molecules (magenta and green) indicated by the magenta and green dashed lines in (H). Bars: 1 μm (D, E), 100 nm (G) and 20 nm (H).

Two-colour imaging is required to study inter-molecular interactions. To investigate the possibility of two-colour imaging, we used McjD-EGFP expressing *E. coli* in which cell membranes were stained

with the red fluorescent probe DiSC₃(5). The DiSC₃(5) dye is cationic and labels putative inner membrane nanodomains, i.e. the fluid lipid network [38], which is made of negatively charged phospholipids such as phosphatidylglycerol. In two-colour cryo-*superSIL* STORM imaging, laser beams of 1.1 kW/cm² 488 nm and 2.2 kW/cm² 642 nm powers were applied in the green and red channels. Two-channel imaging was implemented sequentially with the same camera setting, i.e. 30 ms exposure time per frame, 10 MHz 14 bit EM amplifier readout rate, 5.2× preamp setting and 300 EM gain. Channel alignment was implemented using feature detection during image post-processing. Representative images reveal the distribution of McjD proteins (*green*) and the organization of the fluid lipid network (*red*) (**Fig. 4A**), the latter displaying perceptible helical structures in agreement with previous findings [39]. Furthermore, a high degree of segregation is observed between the McjD and the lipid network. The localization precisions are 10.3 ± 3.8 nm in the green channel and 10.7 ± 3.4 nm in the red channel (**Fig. 4B**).




Fig. 4| *SuperSIL* multi-colour imaging under cryogenic conditions. (A) McjD-EGFP (green) and membrane-DiSC₃(5) (red) in *E.coli* cells imaged in (*top*) wide-field cryo-*superSIL* FM and (*bottom*) cryo-*superSIL* SMLM. The overlays of two-colour images are shown in the right-side column in each row. Scale bar: 1 μ m. (B) The histograms of localization precision of (*top*) McjD-EGFP and (*bottom*) DiSC₃(5).

Similarly to liquid immersion objective lenses, the depth of focus of the combination of the *superSIL* and the dry objective lens is inversely proportional to the square of the effective NA [27]. This results in a narrow depth of focus [40] and thus similar properties to TIRF. We found that the enhanced resolution of the *superSIL* is maintained up to separations of 100 nm from the flat surface of the lens (**Fig. S9-S10**). This paves the way to deploy cryo-fixation for processes in mammalian cells

that require TIRF. We have previously demonstrated TIRF using *superSILs* [21]. To verify that, under cryogenic conditions, the high resolution provided by the ultra-high NA = 2.17 *superSIL* is maintained in the basal periplasmic section adjacent to the lens surface, we imaged plunge-frozen Chinese hamster ovary (CHO) cells. In these the epidermal growth factor receptor (EGFR), a key molecule in cancer research [42], was labelled with Alexa Fluor 488. To extract the areas of highest resolution, we used a standard ‘rolling ball’ background reduction algorithm during image post-processing [43]. The TIRF-like appearance of the image is apparent in **Fig. 5A** [44]. A small cluster of EGFR in the cryo-*superSIL* FM image showed an apparent width of 230 nm (**Fig. 5B, 5C**). Two adjacent clusters 197 nm apart are also clearly distinguishable (**Fig. 5D, 5E**). These values are indistinguishable to those predicted at depths from 10-100 nm (**Fig. S9**).




Fig. 5| Wide-field *superSIL* FM at cryogenic temperature. (A) CHO cells expressing EGFR labelled with Alexa Fluor 488. (B) A cluster of EGFR indicated by the red dashed box in (B). (C) The line profile of the cross-section of the EGFR cluster (red dots) indicated by the red dashed line in (B), and its Gaussian fit (red curve). (D) Two adjacent EGFR clusters indicated by the magenta dashed box in (A). (E) The line profile of the cross-section of the two adjacent EGFR clusters (magenta dots) indicated by the magenta dashed line in (E), and their Gaussian fits (magenta curve). Bars: 2 μm (A) and 0.2 μm (B, D).

It is worth noting that the depth of focus of a *superSIL* microscope depends on the refractive index of the *superSIL* material, allowing a degree of depth “tuning”. This is an important consideration to exploit *superSIL* imaging in CLEM, where lamella thickness is in the range of 50- 300 nm [45]. If required, *superSIL* materials of lower refractive indexes (such as Quartz) can be used to increase the depth of focus approximately 2.2-fold (**Table S1**).

We therefore conclude that by combining a *superSIL* and a low NA dry objective we have achieved our goal of achieving super-resolution STORM in a conventional upright microscope. Indeed, the localization precision and resolution attained are, to our knowledge, the best obtained to date on cell samples. *SuperSIL*-based super-resolution is not limited to STORM, but can be combined with other established super-resolution imaging techniques such as SIM [24] and STED [22]. Importantly, the high NA of the *superSIL* eliminates the barrier to combine cryo-imaging with TIRF, paving the way to the application of vitrification to super-resolution imaging also in mammalian cells.

During our characterization of the *superSIL* setup, we found an unexpected bonus. The ultra-high NA and increased magnification deliver a better resolution also at room temperature than that obtainable from an expensive, specialist STORM system. This means that, even if cryogenic conditions are not required, for example when imaging live cells, a simple add-on based around the use of *superSILs* would enable any non-expert with a basic fluorescence microscope to achieve state-of-the-art super-resolution at low cost. A simple modification of the *superSIL* assembly would easily enable the use of *superSILs* in inverted microscopes too. The method therefore overcomes the costs and complexity inherent to specialist super-resolution set-ups by circumventing the use of expensive objectives, intricate multi-stage illumination paths, specialized sample stages, and high power lasers. A comparison of *superSIL* microscopy vs fluid immersion microscopy is shown in **Table 1**.

Table 1. Characteristics of *superSIL* microscopy versus specialist fluid immersion microscopy.

<i>SuperSIL</i> microscopy	liquid immersion microscopy
NAs in the range of 1.4-2.2, available as required	NA < 1.4
Super-high resolution	High resolution
Super-high light collection efficiency	High light collection efficiency
Suitable for imaging under cryogenic conditions	Not suitable
Low cost	High cost

Given these advantages, we propose that the low-cost, elementary *superSIL* technology has the potential to greatly extend the scope and the reach of super-resolution microscopy in cell biology. The availability to deliver close to 10 nm resolution under cryogenic conditions also has the potential to revolutionize cryo-CLEM, bridging the resolution gap between fluorescence and electron microscopy and allowing sample registration at the nanoscale. Possible schemes for CLEM include imaging of cryogenically-sectioned samples by cryo-*superSIL* super-resolution and transmission EM, or the use of focus ion beam scanning EM to image and produce lamellae suitable for the cryo-*superSIL* microscope. Given the possibility of scaling down *superSILs* to very small and custom-variable sizes,

it is possible to envisage these devices being incorporated into EM systems, allowing true correlative microscopy without the need to move the sample and relocate areas of interest.

In summary, the ability of *superSILs* to achieve a very high NA simply and inexpensively without the use of immersion oil has the potential not only to transform super-resolution microscopy and CLEM, but also to increase the resolution attainable via fluorescence in any non-specialist laboratory at low-cost.

Methods

SuperSIL microscope: The light source was a laser beam combiner (Omicron, LightHUB), including 488 nm and 642 nm laser lines, and a 470 nm collimated LED (Thorlabs, M470L3-C5). For the single colour imaging of McjD-EGFP and PH1735-EGFP *E. coli* cells, the filter sets consisted of a 469/35 nm bandpass filter (Semrock, FF01-469/35-25) as excitation filter, a 484 nm beam splitter (Semrock, FF484-FDi0) as dichroic filter, and a 512/25 nm bandpass filter (Semrock FF01-512/25-25) as emission filter. For the two-colour imaging of McjD-EGFP and membrane- DiSC₃(5) *E. coli* cells, the filter sets consisted of quad-edge dichroic beam splitter (Semrock, Di01-R405/488/543/635-25×36) as dichroic filter and quad-band bandpass filter (Semrock, FF01-446/523/600/677-25) as emission filter, also used in the DNA origami nanoruler measurement.

SuperSIL assembly: The 1 mm diameter *superSILs* (Knight Optical, UK Ltd) were made of cubic zirconia (ZrO₂), the cubic crystalline form of zirconium dioxide (ZrO₂). The lenses have a refractive index of 2.17 and an Abbe number of 33.54 at the wavelength of 512 nm. This provides a high refractive index with medium dispersion suitable for *superSIL* FM. The assemblies were characterized by use of a coordinate measuring machine (OGP SmartScope ZIP 250 Coordinate Measuring Machine) to ensure the angle between the platinum disk and the *superSIL*'s flat surface is < 1°.

SMLM data analysis: ZEISS ZEN software was used to process and render STORM images. In STORM image processing, various peak mask sizes were applied depending on pixel resolution and PSF size in each raw data set. The fit model was a two-dimensional Gaussian fit, and only single emitters from fluorophores were taken into account, whereas all multiple emitters were discarded. Following localization, displacements of molecules from drifts in the reconstructed images were corrected using feature detection and cross correlation. The counts from the raw images were firstly converted to the signal counts by deducting bias offset. Then the signal counts were converted to signal electrons by multiplying the preamp. Finally, the signal electrons were converted to photon numbers by adding the detector electron-multiplying gain.

Sample preparation:

Bacterial cell culture and staining. *E.coli* strain C43 expressing McjD fused with EGFP were prepared for imaging. Briefly, the overnight starter culture was diluted 1:100 in fresh LB media supplemented with 50 µg/ml kanamycin and grown at 37 °C until an optical density at 600 nm (O.D.₆₀₀) of 0.6 was achieved. Then the expression of McjD was induced by adding IPTG (Isopropyl β-D-1-thiogalactopyranoside) to a final concentration of 1 mM at 25 °C overnight. The cells were spun down before freezing. The EGFP counts were measured using a Spectramax microplate reader (Molecular Devices) from 1 ml of culture re-suspended in 200 µl PBS (Thermofisher). For the analysis, the bacterial pellet was re-suspended in 2 ml of PBS to an optical density (O.D.₆₀₀) of ~ 0.6 and kept on ice until plunge-freezing. For two-colour imaging, *E.coli* expressing McjD-EGFP were re-suspended in 250 µl PBS to O.D.₆₀₀ ~ 0.6 and centrifuged at 10000g for 5 min to wash out residues of culture medium. The pellet was then re-suspended in 25 µl of 100 nM DiSC₃(5) in DPBS and incubated for 15 min on ice. 2.5 µl of sample were applied to the flat surface of each *superSIL* immediately prior to plunge-freezing.

Mammalian cell culture and staining. All reagents unless otherwise stated were from Thermo Scientific, UK. CHO cells expressing wtEGFR under an inducible Tet-ON promoter [46] were grown in 5% CO₂ in air at 37 °C in phenol-red free Dulbecco's Modified Eagle medium (DMEM) supplemented with 10% (v/v) fetal bovine serum, 2 mM glutamine, 100 µg/ml hygromycin B and 100 µg/ml geneticin. All cells used were regularly tested for mycoplasma contamination. Cells were seeded at a density of 10⁵/ml on *superSILs* passivated with PEG-BSA nanogel as described previously [47, 48]. Briefly, *superSILs* were etched with Piranha solution for 10 min, and thoroughly rinsed. Priming was performed for 5 min with Vectabond reagent (Vectorlabs) diluted 1:50 in acetone, followed by thorough rinsing. PEG-BSA nanogel was applied for 1h at 37 °C, rinsed twice with PBS, capped with 20 mg/ml BSA in PBS for 1 h 37 °C, quenched with 1 M Tris pH 8.0 for 15 min, followed by three washes in PBS. Cells were cultured for 48 h, rinsed, subjected to nutrient starvation for 2 h at 37 °C to wash out EGFR ligands from the serum and then labelled with 5 nM EGF conjugated to Alexa Fluor 488 (Thermo Scientific) for 30 min at 4 °C. Clustering of EGFR was induced by a 5 min incubation at 37 °C prior to plunge-freezing.

DNA origami nanorulers. Samples were prepared according to manufacturer's instructions. Briefly, *superSILs* were glow discharged for 120 sec (negative) using a Quorum Q150T ES system, then washed 3 times with PBS, then immersed for 5 min in a solution of BSA-biotin 1 mg/ml in PBS and washed a further 3 times with PBS. Coating with neutravidin 1 mg/ml in PBS was also performed by immersion for 5 min, followed by three washes in PBS + 10 mM MgCl₂ (immobilization buffer). 2.5 µl of DNA Origami diluted in immobilization buffer to a final concentration of 1:100 were applied to the flat surface of the *superSILs*, prior to blotting and plunge-freezing.

Plunge freezing. The superSILs were etched for 15 min with Piranha solution (3:1 concentrated sulphuric acid, 30% H₂O₂, both from Sigma-Aldrich) then rinsed with plenty of water and left to air dry for 1 h. SuperSILs were then glow-discharged for 120 s (negative) prior to sample loading and freezing using a Quorum Q150T ES system. Samples were frozen by plunge-freezing using FEI Vitrobot MKIV according to manufacturer's instructions. The chamber was equilibrated to 4 °C, 95% relative humidity. Blotting was performed manually.

Author Contributions

* L.W. and M.L.M conceived and designed the research. B.B. and L.W. constructed the cryo-SIL microscope and implemented the measurements. S.A. and C.S. manufactured the superSIL assembly. L.C.Z. and M.C.D. developed sample plunge freezing protocol. L.C.Z., M.R., A.N.M and S.R.N. prepared biological samples. K.B. supervised M.R. in the bacterial sample preparation. L.W., B.B., D.R., D.T.C and M.L.M analysed data. D.T.C., M.L.M and L.W. wrote the manuscript. All the authors discussed and commented on the manuscript.

ACKNOWLEDGEMENTS

We thank John Collier for his continuous support and contributions to this project and Simon Nestead and David Drew for kindly providing us with the expression construct for PH1735. This work has been funded by Medical Research Council grants (MR/K015591/1) to MMF and (MR/N020103/1) to KB.

References

1. Gustafsson, M.G., *Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy*. Journal of microscopy, 2000. **198**(2): p. 82-87.
2. Hell, S.W. and J. Wichmann, *Breaking the diffraction resolution limit by stimulated emission: stimulated-emission-depletion fluorescence microscopy*. Optics letters, 1994. **19**(11): p. 780-782.
3. Rust, M.J., M. Bates, and X. Zhuang, *Sub-diffraction-limit imaging by stochastic optical reconstruction microscopy (STORM)*. Nature methods, 2006. **3**(10): p. 793.
4. Betzig, E., et al., *Imaging intracellular fluorescent proteins at nanometer resolution*. Science, 2006. **313**(5793): p. 1642-1645.
5. Hess, S.T., T.P. Girirajan, and M.D. Mason, *Ultra-high resolution imaging by fluorescence photoactivation localization microscopy*. Biophysical journal, 2006. **91**(11): p. 4258-4272.
6. Kellenberger, E., et al., *Artefacts and morphological changes during chemical fixation*. Journal of microscopy, 1992. **168**(2): p. 181-201.
7. De Boer, P., J.P. Hoogenboom, and B.N. Giepmans, *Correlated light and electron microscopy: ultrastructure lights up!* Nature methods, 2015. **12**(6): p. 503.
8. Thompson, R.E., D.R. Larson, and W.W. Webb, *Precise nanometer localization analysis for individual fluorescent probes*. Biophysical Journal, 2002. **82**(5): p. 2775-2783.

9. Shtengel, G., et al., *Interferometric fluorescent super-resolution microscopy resolves 3D cellular ultrastructure*. Proceedings of the National Academy of Sciences, 2009. **106**(9): p. 3125-3130.
10. Kaufmann, R., et al., *Super-resolution microscopy using standard fluorescent proteins in intact cells under cryo-conditions*. Nano letters, 2014. **14**(7): p. 4171-4175.
11. Liu, B., et al., *Three-dimensional super-resolution protein localization correlated with vitrified cellular context*. Scientific reports, 2015. **5**: p. 13017.
12. Weisenburger, S., et al., *Cryogenic optical localization provides 3D protein structure data with Angstrom resolution*. Nature methods, 2017. **14**(2): p. 141.
13. Furubayashi, T., et al., *Three-Dimensional Localization of an Individual Fluorescent Molecule with Angstrom Precision*. Journal of the American Chemical Society, 2017. **139**(26): p. 8990-8994.
14. Nahmani, M., et al., *High-numerical-aperture cryogenic light microscopy for increased precision of superresolution reconstructions*. Proceedings of the National Academy of Sciences, 2017. **114**(15): p. 3832-3836.
15. Faoro, R., et al., *Aberration-corrected cryoimmersion light microscopy*. Proceedings of the National Academy of Sciences, 2018: p. 201717282.
16. Mattheyses, A.L., S.M. Simon, and J.Z. Rappoport, *Imaging with total internal reflection fluorescence microscopy for the cell biologist*. J Cell Sci, 2010. **123**(21): p. 3621-3628.
17. Terris, B., et al., *Near - field optical data storage using a solid immersion lens*. Applied Physics Letters, 1994. **65**(4): p. 388-390.
18. Chen, R., et al., *A complete and computationally efficient numerical model of aplanatic solid immersion lens scanning microscope*. Optics express, 2013. **21**(12): p. 14316-14330.
19. Zhang, J., C. See, and M. Somekh, *Imaging performance of widefield solid immersion lens microscopy*. Applied optics, 2007. **46**(20): p. 4202-4208.
20. Wang, L., M.C. Pitter, and M.G. Somekh, *Wide-field high-resolution solid immersion fluorescence microscopy applying an aplanatic solid immersion lens*. Applied Optics, 2010. **49**(31): p. 6160-6169.
21. Wang, L., et al., *Highly confined surface imaging by solid immersion total internal reflection fluorescence microscopy*. Optics Express, 2012. **20**(3): p. 3311-3324.
22. Wildanger, D., et al., *Solid Immersion Facilitates Fluorescence Microscopy with Nanometer Resolution and Sub - Ångström Emitter Localization*. Advanced Materials, 2012. **24**(44).
23. Kim, W.-C., et al., *Investigation on achieving super-resolution by solid immersion lens based STED microscopy*. Optics Express, 2017. **25**(14): p. 16629-16642.
24. Wang, L., M.C. Pitter, and M.G. Somekh, *Wide-field high-resolution structured illumination solid immersion fluorescence microscopy*. Optics Letters, 2011. **36**(15): p. 2794-2796.
25. Liu, Z., et al., *High resolution, high collection efficiency in numerical aperture increasing lens microscopy of individual quantum dots*. Applied Physics Letters, 2005. **87**(7): p. 071905.
26. Rabouw, F.T., et al., *Non-blinking single-photon emitters in silica*. Scientific reports, 2016. **6**: p. 21187.
27. Born, M. and E. Wolf, *Principles of optics: electromagnetic theory of propagation, interference and diffraction of light* 2013: Elsevier.
28. Den Dekker, A. and A. Van den Bos, *Resolution: a survey*. JOSA A, 1997. **14**(3): p. 547-557.
29. Choudhury, H.G., et al., *Structure of an antibacterial peptide ATP-binding cassette transporter in a novel outward occluded state*. Proceedings of the National Academy of Sciences, 2014. **111**(25): p. 9145-9150.
30. Mehmood, S., et al., *Structural and functional basis for lipid synergy on the activity of the antibacterial peptide ABC transporter McjD*. Journal of Biological Chemistry, 2016. **291**(41): p. 21656-21668.
31. Bountra, K., et al., *Structural basis for antibacterial peptide self - immunity by the bacterial ABC transporter McjD*. The EMBO journal, 2017. **36**(20): p. 3062-3079.

32. Romano, M., et al., *Structural Basis for Natural Product Selection and Export by Bacterial ABC Transporters*. ACS Chemical Biology, 2018. **13**(6): p. 1598-1609.
33. Yan, Q., et al., *Mechanistic insights into GLUT1 activation and clustering revealed by super-resolution imaging*. Proceedings of the National Academy of Sciences, 2018.
34. Fitzgerald, J.E., J. Lu, and M.J. Schnitzer, *Estimation theoretic measure of resolution for stochastic localization microscopy*. Physical review letters, 2012. **109**(4): p. 048102.
35. Rothmund, P.W., *Folding DNA to create nanoscale shapes and patterns*. Nature, 2006. **440**(7082): p. 297.
36. Dempsey, G.T., et al., *Evaluation of fluorophores for optimal performance in localization-based super-resolution imaging*. Nature methods, 2011. **8**(12): p. 1027.
37. Beis, K., *Structural basis for the mechanism of ABC transporters*. Biochemical Society Transactions, 2015. **43**(5): p. 889-893.
38. Strahl, H. and J. Errington, *Bacterial membranes: structure, domains, and function*. Annual review of microbiology, 2017. **71**: p. 519-538.
39. Barák, I., et al., *Lipid spirals in Bacillus subtilis and their role in cell division*. Molecular microbiology, 2008. **68**(5): p. 1315-1327.
40. Axelrod, D., *Cell-substrate contacts illuminated by total internal reflection fluorescence*. J Cell Biol, 1981. **89**(1): p. 141-5.
41. Wu, Q., et al., *Realization of numerical aperture 2.0 using a gallium phosphide solid immersion lens*. Applied Physics Letters, 1999. **75**(26): p. 4064-4066.
42. Wee, P. and Z. Wang, *Epidermal growth factor receptor cell proliferation signaling pathways*. Cancers, 2017. **9**(5): p. 52.
43. Sternberg, S.R., *Biomedical image processing*. Computer, 1983. **16**(1): p. 22-34.
44. Mutch, L.J., et al., *Polarised Clathrin - Mediated Endocytosis of EGFR During Chemotactic Invasion*. Traffic, 2014. **15**(6): p. 648-664.
45. Lang, C., et al., *Local thickness and composition analysis of TEM lamellae in the FIB*. Microelectronics Reliability, 2014. **54**(9-10): p. 1790-1793.
46. Macdonald, J.L. and L.J. Pike, *Heterogeneity in EGF-binding affinities arises from negative cooperativity in an aggregating system*. Proceedings of the National Academy of Sciences, 2008. **105**(1): p. 112-117.
47. Tessler, L.A., et al., *Nanogel surface coatings for improved single-molecule imaging substrates*. Journal of the Royal Society Interface, 2011. **8**(63): p. 1400-1408.
48. Zanetti-Domingues, L.C., et al., *A systematic investigation of differential effects of cell culture substrates on the extent of artifacts in single-molecule tracking*. PloS one, 2012. **7**(9): p. e45655.